

Manufacturers of Laboratory Fine Chemicals

Head Office: 403 Cosmos Plaza, (Westside Building), Nr DN Nagar Metro Station, Andheri West- 400053 Mumbai • Contact: +91 91360 93115 Factory: N-226 Near Kumbhavali Naka, Tarapur MIDC, Boisar – 401506. Tel: +91 8451043975/ 76 • Email:sales@pallavchemicals.com

CERTIFICATE OF ANALYSIS

Product Name	:	ORTHOPHOSPHORIC ACID 85% FOR HPLC & SPECTROSCOPY	
Code No.	:	02008	
Batch No.	:	PC01232452	
Mfg. Date		JAN-2023	
Date Of Analysis	:	21/01/2023	
Retest Date	:	DEC-2027	

TEST	SPECIFICATION	RESULT
Appearance	Clear colourless viscous liquid.	Clear colourless viscous liquid.
Assay	Min 85.0%	86.32%
Filter test (1:10 diluted)	In accordance	Complies
Density (at 25°C)	About 1.70 g/ml	1.70 g/ml
Absorption maxima in 1 M solution		-he'
at 210 nm	0.07	0.04
at 220 nm	0.05	0.03
at 230 nm	0.04	0.01
at 250 nm	0.04	0.009
at 500 nm	0.02	0.006

REMARK : This product complies as per the specifications of Pallav Chemicals & Solvents Pvt. Ltd. Manufacturers

This is an electronically generated Certificate and is valid without a signature.

For Issues relating to Quality you may send us a mail on sales@pallavchemicals.com or call on 022 4928 4000

Website: www.pallavchemicals.com