

**Manufacturers of Laboratory Fine Chemicals** 


Head Office: 403 Cosmos Plaza, (Westside Building), Nr DN Nagar Metro Station, Andheri West- 400053 Mumbai • Contact: +91 91360 93115 Factory: N-226 Near Kumbhavali Naka, Tarapur MIDC, Boisar – 401506. Tel: +91 8451043975/ 76 • Email:sales@pallavchemicals.com

## **CERTIFICATE OF ANALYSIS**

<b>Product Name</b>	•	BUFFER SOLUTION PH 4.0 (PHTHALATE)	
Cas No.	••	NA	
Code No.	••	00531	
Batch No.	••	PCS08230116	
Mfg. Date	••	AUG-2023	
Date Of Analysis	••	13/08/2023	
Retest Date	:	JULY-2025	

TEST	SPECIFICATION	RESULT
Appearance	Clear colorless liquid.	Clear colorless liquid.
pH (at 25 °C)	$4.0 \pm 0.05$	4.03

REMARK: This product complies as per the specifications of Pallav Chemicals & Solvents Pvt. Ltd.



For Issues relating to Quality you may send us a mail on sales@pallavchemicals.com or call on 022 4928 4000

Website: www.pallavchemicals.com