



## **CERTIFICATE OF ANALYSIS**

<b>Product Name</b>	:	<b>BUFFER CAPSULES PH 7.0</b>
<b>Cas No.</b>	:	<b>NA</b>
<b>Code No.</b>	:	<b>00541</b>
<b>Batch No.</b>	:	<b>PC09231568</b>
<b>Mfg. Date</b>	:	<b>SEP-2023</b>
<b>Date Of Analysis</b>	:	<b>21/09/2023</b>
<b>Retest Date</b>	:	<b>AUG-2028</b>

<b>TEST</b>	<b>SPECIFICATION</b>	<b>RESULT</b>
Appearance	White Uniform tablets	White Uniform tablets
pH of one tablet dissolved in 100 ml distilled water	7.00 ± 0.05	7.03

REMARK : This product complies as per the specifications of **Pallav Chemicals & Solvents Pvt. Ltd.**

**This is an electronically generated Certificate and is valid without a signature.**

For Issues relating to Quality you may send us a mail on [sales@pallavchemicals.com](mailto:sales@pallavchemicals.com) or call on **022 4928 4000**

Website: [www.pallavchemicals.com](http://www.pallavchemicals.com)