



## **CERTIFICATE OF ANALYSIS**

|                         |   |                                |
|-------------------------|---|--------------------------------|
| <b>Product Name</b>     | : | <b>BUFFER SOLUTION pH 10.0</b> |
| <b>Cas No.</b>          | : | <b>NA</b>                      |
| <b>Code No.</b>         | : | <b>00537</b>                   |
| <b>Batch No.</b>        | : | <b>PCS10230163</b>             |
| <b>Mfg. Date</b>        | : | <b>OCT-2023</b>                |
| <b>Date Of Analysis</b> | : | <b>20/10/2023</b>              |
| <b>Retest Date</b>      | : | <b>SEP-2025</b>                |

| <b>TEST</b>   | <b>SPECIFICATION</b>    | <b>RESULT</b>           |
|---------------|-------------------------|-------------------------|
| Appearance    | Clear colorless liquid. | Clear colorless liquid. |
| pH (at 25 °C) | 10.00 ± 0.05            | 10.03                   |

REMARK : This product complies as per the specifications of **Pallav Chemicals & Solvents Pvt. Ltd.**

**This is an electronically generated Certificate and is valid without a signature.**

For Issues relating to Quality you may send us a mail on **sales@pallavchemicals.com** or call on **022 4928 4000**

Website: [www.pallavchemicals.com](http://www.pallavchemicals.com)