



CERTIFICATE OF ANALYSIS

| | | |
|-------------------------|---|-------------------------------|
| Product Name | : | BUFFER CAPSULES PH 9.2 |
| Cas No. | : | NA |
| Code No. | : | 00542 |
| Batch No. | : | PC08241302 |
| Mfg. Date | : | AUG-2024 |
| Date Of Analysis | : | 12/08/2024 |
| Retest Date | : | JULY-2029 |

| TEST | SPECIFICATION | RESULT |
|--|-----------------------|-----------------------|
| Appearance | White Uniform tablets | White Uniform tablets |
| pH of one tablet dissolved in 100 ml distilled water | 9.20 ± 0.05 | 9.18 |

REMARK : This product complies as per the specifications of **Pallav Chemicals & Solvents Pvt. Ltd.**

This is an electronically generated Certificate and is valid without a signature.

For Issues relating to Quality you may send us a mail on sales@pallavchemicals.com or call on **022 4928 4000**

Website: www.pallavchemicals.com